



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/520,332

09/14/2005

Walter Supper

18477

5111

272 7590 02/27/2009
SCULLY, SCOTT, MURPHY & PRESSER, P.C.
400 GARDEN CITY PLAZA
SUITE 300
GARDEN CITY, NY 11530

EXAMINER

ARAJ, MICHAEL J

ART UNIT

PAPER NUMBER

3775

MAIL DATE

DELIVERY MODE

02/27/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/520,332	Applicant(s) SUPPER ET AL.	
	Examiner MICHAEL J. ARAJ	Art Unit 3775	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 02 October 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1 and 5-16 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1 and 5-16 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Objections

Claims 5 -11 are objected to because of the following informalities: Claim 5 is dependent upon claim 4 which has been cancelled. Claims 6-11 are dependent thereon and are also objected to. Appropriate correction is required.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 5-11 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 5, line 2 recites the limitation "the guide". There is insufficient antecedent basis for this limitation in the claim.

Claim 6, lines 1-2 recites the limitation "the guide". There is insufficient antecedent basis for this limitation in the claim.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 5-7, 12, 15 and 16 are rejected under 35 U.S.C. 102(b) as being anticipated by Hayes et al. (U.S. Patent No. 5,020,519).

Hayes et al. disclose a device comprising a base member (60), comprising a first claw (62), with a distal bearing surface, a second claw having a two-part construction (80 and 14) having a proximal part and a distal part, wherein the distal part is displaceable relative to the proximal part to enable displacement of the second claw in an anteroposterior direction, and wherein displacement of the second claw in a craniocaudal direction is enabled by a parallel displacement device (48). The distal part of the second claw comprises a guide (16), comprising a scale (72), a projection (18) having catches (threads of the guide). The second claw is arranged in such a way that rotation of the second claw relative to the first claw is effectable in a varus-valgus direction, in an internal-external direction in a flexion-extension direction and are effectable independently of one another if one so desired to so.

With regard to the statement of intended use and other functional statements, they do not impose any structural limitations on the claims distinguishable over Daly et al., which is capable of being used as claimed if one so desires to do so. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law of anticipation does not require that the reference “teach” what the subject patent teaches, but rather it is only necessary that the claims under attack “read on” something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not

differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. Ex parte Masham, 2 USPQ2d 1647 (1987).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 8 and 9 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hayes et al. (U.S. Patent No. 5,020,519) in view of Daly et al. (U.S. Patent No. 5,536,271).

Hayes et al. disclose the claimed invention except for a locking device located on the second claw of the device. Daly et al. disclose a locking mechanism to selectively lock the device against movement relative to the other parts of the device (Column 6, lines 7-21). It would have been obvious to one skilled in the art at the time the invention was made to have created the device of Hayes et al. with a locking device on the second claw of the device in view of Daley et al., in order to have better control of the second claws multiple moving parts.

Claims 13 and 14 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hayes et al. (U.S. Patent No. 5,020,519) in view of Niwa et al (U.S. Patent No. 5,213,112).

Hayes et al. disclose the claimed invention except for a force display for the force applied in a craniocaudal direction. Niwa et al. disclose a tensioning meter to monitor the forces being applied between two bones (abstract) in order to produce reproducible results. It would have been obvious to one skilled in the art at the time the invention was made to have created the device of Hayes et al. with a force display in view of Niwa et al., in order to have control of the forces being applied to the joint and to have surgeries that are more easily reproducible.

Allowable Subject Matter

Claims 10 and 11 would be allowable if rewritten to overcome the rejection(s) under 35 U.S.C. 112, 2nd paragraph, set forth in this Office action and to include all of the limitations of the base claim and any intervening claims.

Response to Arguments

Applicant's arguments with respect to claims 1, 5-9 and 13-16 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL J. ARAJ whose telephone number is (571)272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3775

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Michael J Araj/
Examiner, Art Unit 3775

/Eduardo C. Robert/

Supervisory Patent Examiner, Art Unit 3733